PURITAN DATASHEET

Foam Tipped Applicator

Polypropylene Shaft

1605-PSF RECT / 1605-PSFB RECT

NVR

NON-VOLATILE RESIDUE **Test method IDEMA method M7-98**

NVR Solvent - Hexane

NVR 52 µg/swab

The minimum detection limit for NVR is 10 µg/swab

ION CHROMATOGRAPHY TESTING RESULTS

IONIC CLEANLINESS **Test method IDEMA M13-99 (anions) M12-99 (cations)**

ANIONS

Floride	ND	µg/swab
Chloride	4.0000	µg/swab
Nitrate	ND	µg/swab
Bromide	ND	µg/swab
Nitrate	ND	µg/swab
Sulfate	ND	µg/swab
Phosphate	ND	µg/swab

CATIONS

Lithium	ND	µg/swab
Sodium	1.2080	µg/swab
Ammonium	<0.0200	µg/swab
Potassium	0.1000	µg/swab
Magnesium	0.2040	µg/swab
Calcium	0.0920	µg/swab

ND - Non Detected

Test procedures available upon request.

Puritan®
Quality Medical Products Since 1919

Puritan Medical Products Company LLC
Guilford, Maine 04443-0149 USA

Tel 800-321-2313 (US and Canada Only)
207-876-3311

Fax 800-323-4153
207-876-3130

Web www.puritanmedproducts.com

Email sales@puritanmedproducts.com

To Order Call:
Harmony Business Supplies
1-800-899-1255
www.HarmonyCR.com